



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Calibration Laboratory, LLC

3330 East 83rd Place
Merrillville, IN 46410

Fulfills the requirements of

ISO/IEC 17025:2017

and national standards

ANSI/NCSL Z540-1-1994 (R2002)

ANSI/NCSL Z540.3-2006 (R2013)

In the fields of

CALIBRATION and DIMENSIONAL MEASUREMENT

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 15 February 2029
Certificate Number: L2216



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

AND

ANSI/NCSL Z540-1-1994 (R2002)

ANSI/NCSL Z540.3-2006 (R2013)

Calibration Laboratory, LLC

3330 East 83rd Place

Merrillville, IN 46410

Emily Layer 708-596-5800

CALIBRATION AND DIMENSIONAL MEASUREMENT

ISO/IEC 17025 Accreditation Granted: **17 October 2026**

Certificate Number: **L2216** Certificate Expiry Date: **15 February 2029**

CALIBRATION

Chemical Quantities

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
pH Meters ¹ (Fixed Points)	4 pH 7 pH 10 pH	0.016 pH 0.016 pH 0.016 pH	Comparison to Accredited Reference Solutions, RTD Probe

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Measure @ 1 kHz	(0.1 to 10) pF (10 to 100) pF 100 pF to 10 µF (10 to 100) µF (100 to 1 000) µF	4.8 mF/F 0.58 mF/F 0.46 mF/F 0.58 mF/F 4.7 mF/F	Comparison to GenRad 1689M LCR Meter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Capacitance – Source ^{1,15,18} (Fixed Artifacts)	0.1 nF 1 nF 10 nF 100 nF 200 nF 300 nF 400 nF 500 nF	0.58 pF 1.2 pF 12 pF 0.12 nF 0.25 nF 0.37 nF 0.5 nF 0.62 nF	Comparison to Arco SS-32 Standard Capacitors
Capacitance – Source ^{1,15} (Simulation)	(0.19 to 3.3) nF (3.3 to 110) nF (110 to 330) nF (0.33 to 1.1) μ F (1.1 to 3.3) μ F (3.3 to 11) μ F (11 to 33) μ F (33 to 110) μ F (110 to 330) μ F (0.33 to 1.1) mF (1.1 to 3.3) mF (3.3 to 11) mF (11 to 33) mF (33 to 110) mF	4 pF/nF + 7.9 pF 2 pF/nF + 7.8 pF 2 pF/nF + 24 pF 2 nF/ μ F + 0.75 nF 2 nF/ μ F + 0.77 nF 2 nF/ μ F + 0.13 nF 3 nF/ μ F + 2.2 nF 3 nF/ μ F + 78 nF 3 nF/ μ F + 0.23 μ F 3 nF/ μ F + 0.78 μ F 3 μ F/mF + 2.3 μ F 4 μ F/mF + 7.8 μ F 6 μ F/mF + 24 μ F 9 μ F/mF + 77 μ F	Comparison to Fluke 5522A/SC1100 Multi Product Calibrator
AC Current – Source	(9 to 220) μ A (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (0.22 to 2.2) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (2.2 to 22) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.29 nA/ μ A + 23 nA 0.19 nA/ μ A + 14 nA 0.12 nA/ μ A + 10 nA 0.35 nA/ μ A + 15 nA 1.3 nA/ μ A + 80 nA 0.2 μ A/mA + 350 nA 0.19 μ A/mA + 78 nA 0.12 μ A/mA + 47 nA 0.24 μ A/mA + 0.13 μ A 1.3 μ A/mA + 0.8 μ A 0.12 μ A/mA + 4.3 μ A 0.2 μ A/mA + 0.42 μ A 0.12 μ A/mA + 0.44 μ A 0.24 μ A/mA + 0.7 μ A 1.3 μ A/mA + 6 μ A	Comparison to Fluke 5730A Multi Product Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source	(22 to 220) mA (10 to 20) Hz (20 to 40) Hz 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (0.22 to 2.2) A 20 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.25 μ A/mA + 4.3 μ A 0.14 μ A/mA + 8.6 μ A 0.1 μ A/mA + 3.3 μ A 0.2 μ A/mA + 3.6 μ A 1.1 μ A/mA + 10 μ A 0.3 mA/A + 4.4 μ A 0.45 mA/A + 81 μ A 7 mA/A + 0.16 mA	Comparison to Fluke 5730A Multi Product Calibrator
AC Current – Source	(2.2 to 11) A 40 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.46 mA/A + 19 nA 0.95 mA/A + 6.3 μ A 3.6 mA/A + 2.6 μ A	Comparison to Fluke 5730A Multi Product Calibrator, Fluke 5725A Amplifier
AC Current – Source	(0.2 to 2) A (10 to 850) Hz 850 Hz to 6 kHz (6 to 10) kHz (2 to 20) A 10 Hz to 850 Hz 850 Hz to 6 kHz (6 to 10) kHz (20 to 120) A 10 Hz to 850 Hz 850 Hz to 6 kHz (6 to 10) kHz	0.11 mA/A + 36 μ A 0.44 mA/A + 72 μ A 16 mA/A + 62 mA 0.11 mA/A + 36 μ A 0.44 mA/A + 72 μ A 23 mA/A + 94 mA 29 μ A/A + 23 mA 0.44 mA/A + 72 μ A 31 mA/A + 0.7 A	Comparison to Fluke 5730A Multi Product Calibrator, Fluke 52120A Transconductance Amplifier
AC Current – Source ^{1,15}	(30 to 330) μ A (10 to 20) Hz (20 to 45) Hz 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz (10 to 30) kHz	1.6 nA/ μ A + 78 nA 1.2 nA/ μ A + 78 nA 0.97 nA/ μ A + 78 nA 2.3 nA/ μ A + 0.12 μ A 6.2 nA/ μ A + 0.16 μ A 12 nA/ μ A + 0.31 μ A	Comparison to Fluke 5522A/SC1100 Multi Product Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Source ^{1,15}	(0.33 to 3.3) mA		Comparison to Fluke 5522A/SC1100 Multi Product Calibrator
	(10 to 20) Hz	1.5 μ A/mA + 0.18 μ A	
	(20 to 45) Hz	0.97 μ A/mA + 0.12 μ A	
	45 Hz to 1 kHz	0.77 μ A/mA + 0.12 μ A	
	(1 to 5) kHz	1.6 μ A/mA + 0.16 μ A	
	(5 to 10) kHz	3.9 μ A/mA + 0.23 μ A	
	(10 to 30) kHz	7.8 μ A/mA + 0.47 μ A	
	(3.3 to 33) mA		
	(10 to 20) Hz	1.4 μ A/mA + 1.6 μ A	
	(20 to 45) Hz	0.7 μ A/mA + 1.6 μ A	
	45 Hz to 1 kHz	0.31 μ A/mA + 1.6 μ A	
	(1 to 5) kHz	0.62 μ A/mA + 1.6 μ A	
	(5 to 10) kHz	1.6 μ A/mA + 2.3 μ A	
	(10 to 30) kHz	3.1 μ A/mA + 3.1 μ A	
	(33 to 330) mA		
	(10 to 20) Hz	1.4 μ A/mA + 16 μ A	
	(20 to 45) Hz	0.69 μ A/mA + 17 μ A	
	45 Hz to 1 kHz	0.31 μ A/mA + 16 μ A	
	(1 to 5) kHz	0.78 μ A/mA + 39 μ A	
	(5 to 10) kHz	1.6 μ A/mA + 78 μ A	
	10 kHz to 30 kHz	3.1 μ A/mA + 0.16 mA	
	(0.33 to 1.1) A		
	(10 to 45) Hz	1.1 μ A/mA + 78 μ A	
	45 Hz to 1 kHz	0.3 μ A/mA + 78 μ A	
	(1 to 5) kHz	3.6 μ A/mA + 0.78 mA	
	(5 to 10) kHz	15 μ A/mA + 3.9 mA	
	(1.1 to 3) A		
	(10 to 45) Hz	1.4 mA/A + 79 μ A	
	45 Hz to 1 kHz	0.47 mA/A + 78 μ A	
	(1 to 5) kHz	4.7 mA/A + 0.78 mA	
	(5 to 10) kHz	19 mA/A + 3.9 mA	
	(3 to 11) A		
	(45 to 100) Hz	0.47 mA/A + 1.6 mA	
	100 Hz to 1 kHz	0.78 mA/A + 1.6 mA	
	(1 to 5) kHz	23 mA/A + 1.6 mA	
	(11 to 20) A		
	(45 to 100) Hz	0.93 mA/A + 3.9 mA	
	100 Hz to 1 kHz	1.2 mA/A + 3.9 mA	
	(1 to 5) kHz	23 mA/A + 3.9 mA	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current Source – Current Clamps ^{1,15}	(45 to 65) Hz (10 to 16.5) A (16.5 to 150) A (150 to 1 025) A (65 to 440) Hz (10 to 16.5) A (16.5 to 150) A (150 to 1 025) A	2.2 mA/A + 3.5 mA 2.1 mA/A + 44 mA 2.5 mA/A + 52 mA 4.6 mA/A + 30 mA 6.2 mA/A + 33 mA 6.4 mA/A + 62 mA	Comparison to Fluke 5522A/SC1100 Multi Product Calibrator, Fluke 50-Turn Coil
AC Current Source – Current Clamps	(0.2 to 25) A (6 kHz to 10 kHz) (0.2 to 50) A (3 kHz to 6 kHz) (0.2 to 50) A (1 kHz to 3 kHz) (0.2 to 50) A 10 Hz to 1 kHz (50 to 75) A (3 kHz to 6 kHz) (50 to 300) A (1 kHz to 3 kHz) (50 to 500) A 10 Hz to 1 kHz (500 to 1 000) A 300 Hz to 1 kHz (500 to 3 000) A 10 Hz to 300 Hz	39 mA/A + 16 mA 12 mA/A + 16 mA 6.2 mA/A + 16 mA 5.4 mA/A + 11 mA 12 mA/A + 0.15 A 6.2 mA/A + 0.16 A 5.4 mA/A + 0.11 A 3.6 mA/A + 0.45 A 3.6 mA/A + 0.45 A	Comparison to Fluke 5730A Multi Product Calibrator, Fluke 52120A Transconductance Amplifier, Fluke 3KA 25-Turn Coil
DC Current – Source ^{1,15}	(0 to 330) μ A (0.33 to 3.3) mA (3.3 to 33) mA (33 to 330) mA (0.33 to 1.1) A (1.1 to 3) A (3 to 11) A (11 to 20) A	0.11 nA/ μ A + 17 nA 78 nA/mA + 39 nA 77 nA/mA + 37 nA 77 nA/mA + 2.2 μ A 0.13 mA/A + 75 μ A 0.29 mA/A + 32 μ A 0.39 mA/A + 0.41 mA 0.78 mA/A + 0.59 mA	Comparison to Fluke 5522A/SC1100 Multi Product Calibrator
DC Current – Source	(0 to 220) μ A (0.22 to 2.2) mA (2.2 to 22) mA (22 to 220) mA (0.22 to 2.2) A	50 μ A/A + 7 nA 35 μ A/A + 7 nA 35 μ A/A + 40 nA 45 μ A/A + 0.7 μ A 80 μ A/A + 12 μ A	Comparison to Fluke 5730A Multi Product Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC Current – Source	(2.2 to 11) A	0.36 mA/A + 0.51 mA	Comparison to Fluke 5730A Multi Product Calibrator, Fluke 5725A Amplifier
DC Current – Source	(0 to 100) A	35 μ A/A + 23 mA	Comparison to Fluke 5730A Multi Product Calibrator, Fluke 52120A Transconductance Amplifier
DC Current Source – Current Clamps ^{1,15}	(10 to 150) A (150 to 1 025) A	2 mA/A + 1.4 mA 2 mA/A + 48 mA	Comparison to Fluke 5522A/SC1100 Multi Product Calibrator, Fluke 50-Turn Coil
DC Current Source – Current Clamps	(0 to 2 500) A	5.7 mA/A + 78 mA	Comparison to Fluke 5730A Multi Product Calibrator, Fluke 52120A Transconductance Amplifier, Fluke 3KA 25-Turn Coil
DC Current – Measure ¹⁸	10 pA to 10 nA (20 to 30) A	16 pA/nA + 3.4 pA 0.78 mA/A + 4.5 mA	Comparison to Transmille 8081 8.5 Digit Multimeter
DC Current – Measure ^{1,15}	(0 to 200) μ A (0.2 to 2) mA (2 to 20) mA (20 to 200) mA (0.2 to 2) A	0.33 mA/A + 12 nA 0.35 mA/A + 69 nA 2.8 μ A/A + 9.5 μ A 0.37 mA/A + 7.5 μ A 0.74 mA/A + 0.11 mA	Comparison to Keithley 2002 8.5 Digit Multimeter
DC Current – Measure ^{1,15}	(1 to 3) A	1.23 mA/A + 0.63 mA	Comparison to Agilent 34401A 6.5 Digit Multimeter
DC Current – Measure ¹⁸	(0 to 200) μ A 200 μ A to 2 mA (2 to 20) mA (20 to 200) mA 200 mA to 2 A (2 to 20) A	13 pA/ μ A + 0.3 nA 13 nA/mA + 3 nA 16 nA/mA + 29 nA 44 nA/mA + 0.55 μ A 0.17 mA/A + 12 μ A 0.36 mA/A + 0.44 mA	Comparison to Fluke 8508A 8.5 Digit Multimeter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure ¹⁸	0.1 nA to 0.1 mA (10 to 40) Hz 40 Hz to 1 kHz (1 to 10) kHz (20 to 30) A (10 to 40) Hz 40 Hz to 1 kHz	0.92 nA/μA + 16 nA 0.51 nA/μA + 13 nA 1.3 nA/μA + 31 nA 1.7 mA/A + 12 mA 1.3 mA/A + 9.2 mA	Comparison to Transmille 8081 8.5 Digit Multimeter
AC Current – Measure ¹⁸	(20 to 200) μA 10 Hz to 10 kHz (10 to 30) kHz (0.2 to 2) mA 10 Hz to 10 kHz (10 to 30) kHz (2 to 20) mA 10 Hz to 10 kHz (10 to 30) kHz (20 to 200) mA 10 Hz to 10 kHz (10 to 30) kHz (0.2 to 2) A 10 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (2 to 20) A 10 Hz to 2 kHz (2 to 10) kHz	0.57 nA/μA + 18 nA 0.6 nA/μA + 19 nA 0.28 μA/mA + 0.19 μA 0.73 μA/mA + 0.18 μA 0.27 μA/mA + 1.9 μA 0.68 μA/mA + 1.8 μA 0.25 μA/mA + 19 μA 0.62 μA/mA + 18 μA 0.56 mA/A + 0.19 mA 0.92 mA/A - 0.84 μA 2.4 mA/A + 0.19 mA 0.72 mA/A + 1.9 mA 2 mA/A + 1.8 mA	Comparison to Fluke 8508A 8.5 Digit Multimeter
AC Current – Measure	(20 to 200) mA (1 to 10) Hz 10 Hz to 10 kHz (10 to 30) kHz (0.2 to 2) A 10 Hz to 2 kHz (2 to 10) kHz (10 to 30) kHz (2 to 20) A 10 Hz to 2 kHz (2 to 10) kHz	0.29 μA/mA + 19 μA 0.24 μA/mA + 19 μA 0.54 μA/mA + 19 μA 0.55 mA/A + 0.19 mA 0.63 mA/A + 0.19 mA 2.3 mA/A + 0.19 mA 0.70 mA/A + 1.9 mA 1.9 mA/A + 1.9 mA	Comparison to Fluke 8508A 8.5 Digit Multimeter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure ^{1,15}	Up to 200 μ A (20 to 50) Hz (50 to 200) Hz 200 Hz to 1 kHz (1 to 10) kHz (0.2 to 2) mA (20 to 50) Hz (50 to 200) Hz 200 Hz to 1 kHz (1 to 10) kHz (10 to 30) kHz (30 to 50) kHz (50 to 100) kHz	3.5 mA/A + 36 nA 2 mA/A + 38 nA 4 mA/A + 35 nA 4 mA/A + 0.27 μ A 3.5 mA/A + 0.34 μ A 1.5 mA/A + 0.3 μ A 1.2 mA/A + 0.37 μ A 0.51 mA/A + 3.2 μ A 1.6 mA/A + 3.1 μ A 2.1 mA/A + 3 μ A 5 mA/A + 0.3 μ A	Comparison to Keithley 2002 8.5 Digit Multimeter
	(2 to 20) mA (20 to 50) Hz (50 to 200) Hz 200 Hz to 1 kHz (1 to 10) kHz (10 to 30) kHz (30 to 50) kHz (50 to 100) kHz (20 to 200) mA (20 to 50) Hz (50 to 200) Hz 200 Hz to 1 kHz (1 to 10) kHz (10 to 30) kHz (30 to 50) kHz (50 to 100) kHz	3 mA/A + 3.4 μ A 1.5 mA/A + 3.6 μ A 1.2 mA/A + 3.7 μ A 0.54 mA/A + 30 μ A 1.6 mA/A + 28 μ A 2.1 mA/A + 28 μ A 4.1 mA/A + 25 μ A 3 mA/A + 33 μ A 1.5 mA/A + 34 μ A 1.2 mA/A + 35 μ A 0.94 mA/A + 0.21 mA 4.4 mA/A + 0.17 mA 9.6 mA/A + 0.13 mA 29.8 mA/A + 71 μ A	
AC Current – Measure ^{1,15}	(0.2 to 2) A (20 to 50) Hz (50 to 200) Hz 200 Hz to 1 kHz (1 to 10) kHz	4.1 mA/A + 0.35 mA 2.3 mA/A + 0.35 mA 3.5 mA/A + 0.35 mA 5.2 mA/A + 0.36 mA	Comparison to Keithley 2002 8.5 Digit Multimeter
AC Current – Measure ^{1,15}	(1 to 3) A (3 to 5) Hz (5 to 10) Hz 10 Hz to 5 kHz	11 mA/A + 1.8 mA 3.5 mA/A + 1.9 mA 1.5 mA/A + 2 mA	Comparison to Agilent 34401A 6.5 Digit Multimeter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Current – Measure ^{1, 18}	60 Hz (5 to 3 000) A	13 mA/A – 89 mA	Comparison to AEMC 30K-24-2 Current Probe, Agilent 34401A 6.5 Digit Multimeter
Inductance – Measure	1 kHz (10 to 100) μ H (0.1 to 1) mH (1 to 10) mH (10 to 100) mH (0.1 to 1) H (1 to 10) H	0.1 nH/ μ H + 0.2 μ H 0.6 μ H/mH + 0.2 μ H 0.8 μ H/mH + 0.2 μ H 0.8 μ H/mH + 2 μ H 0.8 mH/H + 20 μ H 0.8 mH/H + 0.2 mH	Comparison to GenRad 1689M LCR Meter
Inductance – Source ^{1,15,18} (Fixed Artifacts)	100 μ H 500 μ H 1 mH 10 mH 50 mH 100 mH 200 mH 1 H 2 H 10 H	0.36 μ H 0.62 μ H 1.4 μ H 14 μ H 73 μ H 0.14 mH 0.24 mH 1.5 mH 2.7 mH 14 mH	Comparison to GenRad 1482 Series Standard Inductors
Phase – Measure/Source ^{1,15} (0.01 to 100) V	(0 to 360)° 1 Hz to 100 Hz 1 Hz to 1 kHz (1 to 10) kHz (10 to 100) kHz 100 kHz to 1 MHz	0.037° 0.056° 0.068° 0.13° 0.36°	Comparison to Keysight 53220A Phase Meter, Fluke 5522A/SC1100 Multi Product Calibrator
DC Power – Source ^{1,15,17}	(0 to 336) W (336 to 3 060) W (3 060 to 20 910) W	0.037 % of reading 0.053 % of reading 0.12 % of reading	Comparison to Fluke 5522A/SC1100 Multi Product Calibrator
AC Power – Source ^{1,15,17}	(45 to 65) Hz (0.11 to 3) mW (3 to 11) mW (11 to 30) mW (30 to 110) mW (110 to 300) mW (300 to 730) mW	0.11 % of reading 0.12 % of reading 0.17 % of reading 0.12 % of reading 0.16 % of reading 0.13 % of reading	Comparison to Fluke 5522A/SC1100 Multi Product Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Power – Source ^{1,15,17}	(45 to 65) Hz (0.73 to 1.5) W (1.5 to 6.8) W (6.8 to 9.2) W (9.2 to 34) W (34 to 92) W (92 to 337) W (337 to 918) W (918 to 2 244) W (2 244 to 4 590) W (4 590 to 11 220) W	0.15 % of reading 0.14 % of reading 0.14 % of reading 0.1 % of reading 0.14 % of reading 0.1 % of reading 0.13 % of reading 0.11 % of reading 0.14 % of reading 0.12 % of reading	Comparison to Fluke 5522A/SC1100 Multi Product Calibrator
Resistance – Source (Fixed Artifacts)	25 Ω 100 Ω 200 Ω 400 Ω 1 k Ω 1.9 k Ω 4 k Ω 10 k Ω 40 k Ω 90 k Ω 100 k Ω 900 k Ω 1 M Ω	0.3 m Ω 1.2 m Ω 2.3 m Ω 4.7 m Ω 12 m Ω 22 m Ω 48 m Ω 0.13 Ω 0.46 Ω 1.1 Ω 1.2 Ω 11 Ω 12 Ω	Comparison to Vishay Resistance Standards
Resistance – Source (Fixed Points)	0 Ω 1 Ω 1.9 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1 k Ω	40 $\mu\Omega$ 95 $\mu\Omega$ 0.18 m Ω 0.23 m Ω 0.44 m Ω 1.5 m Ω 1.9 m Ω 6.5 m Ω	Comparison to Fluke 5730A Multi Product Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Source (Fixed Points)	1.9 k Ω 10 k Ω 19 k Ω 100 k Ω 190 k Ω 1 M Ω 1.9 M Ω 10 M Ω 19 M Ω 100 M Ω	12 m Ω 66 m Ω 0.12 Ω 0.85 Ω 1.6 Ω 13 Ω 39 Ω 0.42 k Ω 1.2 k Ω 10 k Ω	Comparison to Fluke 5730A Multi Product Calibrator
Resistance – Source (Fixed Artifacts)	1 G Ω 10 G Ω	49 k Ω 0.64 M Ω	Comparison to Measurements International Resistance Standards
Resistance – Source ^{1,15} (Simulation)	(0 to 11) Ω (11 to 33) Ω (33 to 110) Ω (110 to 330) Ω (0.33 to 1.1) k Ω (1.1 to 3.3) k Ω (3.3 to 11) k Ω (11 to 33) k Ω (33 to 110) k Ω (110 to 330) k Ω (0.33 to 1.1) M Ω (1.1 to 3.3) M Ω (3.3 to 11) M Ω (11 to 33) M Ω (33 to 110) M Ω (110 to 330) M Ω (0.33 to 1.1) G Ω	31 $\mu\Omega/\Omega$ + 0.8 m Ω 23 $\mu\Omega/\Omega$ + 1.2 m Ω 22 $\mu\Omega/\Omega$ + 1.1 m Ω 22 $\mu\Omega/\Omega$ + 1.7 m Ω 22 $\mu\Omega/\Omega$ + 1.6 m Ω 22 $\mu\Omega/\Omega$ + 16 m Ω 20 $\mu\Omega/\Omega$ + 16 m Ω 20 $\mu\Omega/\Omega$ + 0.16 Ω 22 $\mu\Omega/\Omega$ + 0.17 Ω 25 $\mu\Omega/\Omega$ + 1.6 Ω 46 $\mu\Omega/\Omega$ + 24 Ω 44 $\mu\Omega/\Omega$ + 35 Ω 93 $\mu\Omega/\Omega$ + 160 Ω 0.18 m Ω/Ω + 2.7 k Ω 0.38 m Ω/Ω + 3.9 k Ω 2.3 m Ω/Ω + 84 k Ω 13 m Ω/Ω – 0.18 M Ω	Comparison to Fluke 5522A/SC1100 Multi Product Calibrator
Resistance – Measure ^{1,15}	(0 to 20) Ω (20 to 200) Ω (0.2 to 2) k Ω (2 to 20) k Ω (20 to 200) k Ω (0.2 to 2) M Ω (2 to 20) M Ω (20 to 200) M Ω	9.1 $\mu\Omega/\Omega$ + 0.45 m Ω 13.3 $\mu\Omega/\Omega$ + 1.1 m Ω 5.5 $\mu\Omega/\Omega$ + 11 m Ω 4.8 $\mu\Omega/\Omega$ + 0.12 Ω 30.7 $\mu\Omega/\Omega$ + 1.2 Ω 55 $\mu\Omega/\Omega$ + 0.11 k Ω 0.23 m Ω/Ω + 0.58 k Ω 0.53 m Ω/Ω + 4.7 k Ω	Comparison to Keithley 2002 8.5 Digit Multimeter

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Resistance – Measure ¹⁸	(0.1 to 2) Ω (2 to 20) Ω (20 to 200) Ω (0.2 to 2) k Ω (2 to 20) k Ω (20 to 200) k Ω (0.2 to 2) M Ω (2 to 20) M Ω	30 $\mu\Omega/\Omega$ + 3.6 $\mu\Omega$ 11 $\mu\Omega/\Omega$ + 15 $\mu\Omega$ 8.2 $\mu\Omega/\Omega$ + 47 $\mu\Omega$ 9.6 $\mu\Omega/\Omega$ + 0.48 m Ω 8.8 m $\Omega/\text{k}\Omega$ + 5 m Ω 13 m $\Omega/\text{k}\Omega$ + 33 m Ω 14 $\Omega/\text{M}\Omega$ + 0.75 Ω 24 $\Omega/\text{M}\Omega$ + 88 Ω	Comparison to Fluke 8508A 8.5 Digit Multimeter
Resistance – Measure ¹⁸	(2 to 20) M Ω (20 to 200) M Ω (0.2 to 2) G Ω (2 to 20) G Ω	20 $\Omega/\text{M}\Omega$ + 7.8 Ω 110 $\Omega/\text{M}\Omega$ + 0.67 k Ω 0.22 M $\Omega/\text{G}\Omega$ + 90 k Ω 1.5 M $\Omega/\text{G}\Omega$ + 9.3 M Ω	Comparison to Fluke 8508A 8.5 Digit Multimeter (HV mode)
Resistance – Measure ¹⁸	24 G Ω to 2 T Ω	28 M $\Omega/\text{G}\Omega$ + 30 M Ω	Comparison to Transmille 8081 8.5 Digit Multimeter
Electrical Simulation of RTD Indicating Devices – Source ^{1,15}	Pt 385, 100 Ω (-200 to -80) $^{\circ}\text{C}$ (-80 to 0) $^{\circ}\text{C}$ (0 to 100) $^{\circ}\text{C}$ (100 to 300) $^{\circ}\text{C}$ (300 to 400) $^{\circ}\text{C}$ (400 to 630) $^{\circ}\text{C}$ (630 to 800) $^{\circ}\text{C}$ Pt 385, 200 Ω (-200 to -80) $^{\circ}\text{C}$ (-80 to 0) $^{\circ}\text{C}$ (0 to 100) $^{\circ}\text{C}$ (100 to 260) $^{\circ}\text{C}$ (260 to 300) $^{\circ}\text{C}$ (300 to 400) $^{\circ}\text{C}$ (400 to 600) $^{\circ}\text{C}$ (600 to 630) $^{\circ}\text{C}$ Pt 385, 500 Ω (-200 to -80) $^{\circ}\text{C}$ (-80 to 0) $^{\circ}\text{C}$ (0 to 100) $^{\circ}\text{C}$ (100 to 260) $^{\circ}\text{C}$ (260 to 300) $^{\circ}\text{C}$ (300 to 400) $^{\circ}\text{C}$ (400 to 600) $^{\circ}\text{C}$ (600 to 630) $^{\circ}\text{C}$	0.041 $^{\circ}\text{C}$ 0.041 $^{\circ}\text{C}$ 0.056 $^{\circ}\text{C}$ 0.071 $^{\circ}\text{C}$ 0.079 $^{\circ}\text{C}$ 0.094 $^{\circ}\text{C}$ 0.18 $^{\circ}\text{C}$ 0.034 $^{\circ}\text{C}$ 0.027 $^{\circ}\text{C}$ 0.034 $^{\circ}\text{C}$ 0.034 $^{\circ}\text{C}$ 0.086 $^{\circ}\text{C}$ 0.095 $^{\circ}\text{C}$ 0.095 $^{\circ}\text{C}$ 0.11 $^{\circ}\text{C}$ 0.034 $^{\circ}\text{C}$ 0.041 $^{\circ}\text{C}$ 0.041 $^{\circ}\text{C}$ 0.049 $^{\circ}\text{C}$ 0.064 $^{\circ}\text{C}$ 0.064 $^{\circ}\text{C}$ 0.071 $^{\circ}\text{C}$ 0.086 $^{\circ}\text{C}$	Comparison to Fluke 5522A/SC1100 Multi Product Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of RTD Indicating Devices – Source ^{1,15}	Pt 385, 1 k Ω		Comparison to Fluke 5522A/SC1100 Multi Product Calibrator
	(-200 to -80) °C	0.027 °C	
	(-80 to 0) °C	0.027 °C	
	(0 to 100) °C	0.034 °C	
	(100 to 260) °C	0.041 °C	
	(260 to 300) °C	0.048 °C	
	(300 to 400) °C	0.056 °C	
	(400 to 600) °C	0.056 °C	
	(600 to 630) °C	0.18 °C	
	Pt 3916, 100 Ω		
	(-200 to -190) °C	0.19 °C	
	(-190 to -80) °C	0.034 °C	
	(-80 to 0) °C	0.041 °C	
	(0 to 100) °C	0.048 °C	
	(100 to 260) °C	0.056 °C	
	(260 to 300) °C	0.063 °C	
	(300 to 400) °C	0.071 °C	
	(400 to 600) °C	0.079 °C	
	(600 to 630) °C	0.18 °C	
	Pt 3926, 100 Ω		
	(-200 to -80) °C	0.041 °C	
	(-80 to 0) °C	0.041 °C	
	(0 to 100) °C	0.056 °C	
	(100 to 300) °C	0.071 °C	
	(300 to 400) °C	0.079 °C	
	(400 to 630) °C	0.094 °C	
	Ni 385, 120 Ω		
	(-80 to 0) °C	0.063 °C	
	(0 to 100) °C	0.063 °C	
	(100 to 260) °C	0.11 °C	
	Cu 427, 10 Ω		
	(-100 to 260) °C	0.23 °C	
DC Voltage – Source	(0 to 220) mV	8.2 μ V/V + 0.41 μ V	Comparison to Fluke 5730A Multi Product Calibrator
	(0.22 to 2.2) V	5 μ V/V + 0.77 μ V	
	(2.2 to 11) V	0.45 μ V/V + 3.5 μ V	
	(11 to 22) V	3.5 μ V/V + 0.59 μ V	
	(22 to 220) V	5.8 μ V/V + 51 μ V	
	(220 to 1 100) V	6.5 μ V/V + 0.42 mV	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
DC High Voltage – Source ^{1,15,18}	(10 to 1 400) V (1.4 to 10) kV	0.35 mV/V + 41 mV 0.4 V/kV + 38 mV	Comparison to HV Output monitored with Vitretek 4700 High Voltage Meter, HVL-35 Probe, HVL-100 Probe
	(1 to 20) kV	0.42 V/kV + 62 mV	
	(20 to 30) kV	0.9 V/kV – 9.5 V	
	(30 to 35) kV	1.3 V/kV – 20 V	
	(10 to 40) kV	0.72 V/kV – 0.5 V	
	(40 to 60) kV	1.3 V/kV – 22 V	
	(60 to 80) kV (80 to 100) kV	2.2 V/kV – 77 V 4 V/kV – 220 V	
DC Voltage – Source ^{1,15}	(0 to 330) mV (0.33 to 3.3) V (3.3 to 33) V (33 to 330) V (330 to 1 000) V	13 µV/V + 2 µV 8 µV/V + 4.1 µV 8.9 µV/V + 34 µV 13.6 µV/V + 0.26 mV 14 µV/V + 1.1 mV	Fluke 5522A/SC1100 Multi Product Calibrator
	(0.22 to 2.2) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.23 µV/mV + 3.9 µV 410 nV/mV + 3.8 µV 81 nV/mV + 3.9 µV 0.19 µV/mV + 3.9 µV 0.46 µV/mV + 4.7 µV 1 µV/mV + 9.3 µV 1.3 µV/mV + 19 µV 2.6 µV/mV + 19 µV	
	(2.2 to 22) mV (10 to 20) Hz (20 to 40) Hz 40 Hz to 20 kHz (20 to 50) kHz (50 to 100) kHz (100 to 300) kHz (300 to 500) kHz 500 kHz to 1 MHz	0.23 µV/mV + 3.9 µV 89 nV/mV + 3.9 µV 77 nV/mV + 3.9 µV 0.19 µV/mV + 3.9 µV 0.47 µV/mV + 4.7 µV 1 µV/mV + 9.3 µV 1 µV/mV + 9.3 µV 2.6 µV/mV + 19 µV	
AC Voltage – Source			Comparison to Fluke 5730A Multi Product Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source	(22 to 220) mV		Comparison to Fluke 5730A Multi Product Calibrator
	(10 to 20) Hz	0.23 μ V/mV + 12 μ V	
	(20 to 40) Hz	89 nV/mV + 6.3 μ V	
	40 Hz to 20 kHz	54 nV/mV + 6.2 μ V	
	(20 to 50) kHz	0.12 μ V/mV + 6.2 μ V	
	(50 to 100) kHz	0.31 μ V/mV + 16 μ V	
	(100 to 300) kHz	1 μ V/mV + 9.3 μ V	
	(300 to 500) kHz	1.3 μ V/mV + 27 μ V	
	500 kHz to 1 MHz	2.6 μ V/mV + 47 μ V	
	(0.22 to 2.2) V		
	(10 to 20) Hz	0.23 mV/V + 45 μ V	
	(20 to 40) Hz	85 μ V/V + 16 μ V	
	40 Hz to 20 kHz	37 μ V/V + 8.5 μ V	
	(20 to 50) kHz	62 μ V/V + 9.7 μ V	
	(50 to 100) kHz	78 μ V /V + 31 μ V	
	(100 to 300) kHz	0.31 mV/V + 78 μ V	
	(300 to 500) kHz	0.93 mV/V + 0.2 mV	
	500 kHz to 1 MHz	1.6 mV/V + 0.31 mV	
	(2.2 to 22) V		
	(10 to 20) Hz	0.23 mV/V + 0.46 mV	
	(20 to 40) Hz	85 μ V /V + 0.13 mV	
	40 Hz to 20 kHz	37 μ V/V + 58 μ V	
	(20 to 50) kHz	62 μ V/V + 0.95 mV	
	(50 to 100) kHz	77 μ V/V + 0.2 mV	
	(100 to 300) kHz	0.23 mV/V + 0.62 mV	
	(300 to 500) kHz	0.93 mV/V + 1.9 mV	
	500 kHz to 1 MHz	1.4 mV/V + 3.1 mV	
	(22 to 220) V		
	(10 to 20) Hz	0.23 mV/V + 4.2 mV	
	(20 to 40) Hz	85 μ V/V + 1.7 mV	
	40 Hz to 20 kHz	50 μ V/V + 0.69 mV	
	(20 to 50) kHz	77 μ V /V + 0.96 mV	
	(50 to 100) kHz	0.14 mV/V + 2.4 mV	
	(100 to 300) kHz	0.85 mV/V + 16 mV	
	(220 to 1 100) V		
	(15 to 50) Hz	0.28 mV/V + 16 mV	
	50 Hz to 1 kHz	66 μ V /V + 3.6 mV	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source	(220 to 750) V (30 to 50) kHz (50 to 100) kHz (220 to 1 100) V (1 to 20) kHz (20 to 30) kHz	8.6 mV + 0.47 mV/V 35 mV + 1.8 mV/V 4.7 mV + 0.13 mV/V 8.5 mV + 0.47 mV/V	Comparison to Fluke 5730A Multi Product Calibrator, Fluke 5725A Amplifier
AC Voltage – Source ^{1,15}	(1 to 33) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz (33 to 330) mV (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz (0.33 to 3.3) V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz (100 to 500) kHz (3.3 to 33) V (10 to 45) Hz 45 Hz to 10 kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.62 µV/mV + 5 µV 0.11 µV/mV + 5.1 µV 0.13 µV/mV + 1.3 µV 0.77 µV/mV + 4.9 µV 2.7 µV/mV + 9.4 µV 6.2 µV/mV + 39 µV 0.23 µV/mV + 6.4 µV 0.11 µV/mV + 6.4 µV 0.12 µV/mV + 6.5 µV 0.27 µV/mV + 6.4 µV 0.62 µV/mV + 25 µV 1.6 µV/mV + 54 µV 0.12 mV/V + 47 µV 0.12 mV/V + 47 µV 0.15 mV/V + 47 µV 0.23 mV/V + 39 µV 0.54 mV/V + 97 µV 1.9 mV/V + 0.47 mV 0.23 mV/V + 0.5 mV 0.12 mV/V + 0.47 mV 0.23 mV/V + 0.51 mV 0.27 mV/V + 0.47 mV 0.7 mV/V + 1.2 mV	Comparison to Fluke 5522A/SC1100 Multi Product Calibrator

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Source ^{1,15}	(33 to 330) V 45 Hz to 1 kHz (1 to 10) kHz (10 to 20) kHz (20 to 50) kHz (50 to 100) kHz	0.15 mV/V + 1.7 mV 0.16 mV/V + 4.8 mV 0.19 mV/V + 4.8 mV 0.23 mV/V + 4.9 mV 1.6 mV/V + 39 mV	Comparison to Fluke 5522A/SC1100 Multi Product Calibrator
	(330 to 1 000) V 45 Hz to 1 kHz (1 to 5) kHz (5 to 10) kHz	0.23 mV/V + 7.9 mV 0.19 mV/V + 8.1 mV 0.23 mV/V + 7.9 mV	
AC High Voltage – Source ^{1,15,18}	60 Hz 10 mV to 1 000 V (1 to 10) kV	1.4 mV/V + 0.13 V 1.5 mV/kV + 0.4 V	Comparison to HV Output Monitored with Vitretek 4700 High Voltage Meter
DC Voltage – Measure ^{1,15}	(0 to 200) mV (0.2 to 2) V (2 to 20) V (20 to 200) V (200 to 1 000) V	16.6 μV/V + 2.6 μV 7.3 μV/V + 9.4 μV 7.8 μV/V + 59 μV 19.3 μV/V + 1.1 mV 18.6 μV/V + 4.5 mV	Comparison to Keithley 2002 8.5 Digit Multimeter
DC Voltage – Measure ¹⁸	(0 to 200) mV (0.2 to 2) V (2 to 20) V (20 to 200) V (200 to 1 000) V	6.1 nV/mV + 97 nV 3.5 μV/V + 0.44 μV 3.4 μV/V + 4.1 μV 5.4 μV/V + 38 μV 5.9 μV/V + 0.37 mV	Comparison to Fluke 8508A 8.5 Digit Multimeter
DC High Voltage – Measure ^{1,15,18}	(10 to 1 400) V (1.4 to 10) kV (1 to 20 kV) (20 to 30) kV (30 to 35) kV (10 to 40) kV (40 to 60) kV (60 to 80) kV (80 to 100) kV	0.35 mV/V + 41 mV 0.4 V/kV + 38 mV 0.42 V/kV + 62 mV 0.9 V/kV – 9.5 V 1.3 V/kV – 20 V 0.72 V/kV – 0.5 V 1.3 V/kV – 22 V 2.2 V/kV – 77 V 4 V/kV – 0.22 kV	Comparison to Vitretek 4700 High Voltage Meter, HVL-35 Probe, HVL-100 Probe

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure ^{1,15}	(2 to 200) mV		Comparison to Keithley 2002 8.5 Digit Multimeter
	(10 to 40) Hz	0.19 μ V/mV + 27 μ V	
	(40 to 100) Hz	0.16 μ V/mV + 3.4 μ V	
	100 Hz to 2 kHz	0.16 μ V/mV + 1.4 μ V	
	(2 to 10) kHz	0.31 μ V/mV + 66 nV	
	(10 to 30) kHz	0.3 μ V/mV + 7.5 μ V	
	(30 to 100) kHz	0.76 μ V/mV + 17 μ V	
	(0.2 to 2) V		
	(1 to 10) Hz	0.9 mV/V + 0.3 mV	
	(10 to 50) Hz	0.4 mV/V + 0.3 mV	
	(50 to 100) Hz	0.25 mV/V + 0.31 mV	
	100 Hz to 2 kHz	0.2 mV/V + 0.22 mV	
	(2 to 10) kHz	0.19 mV/V + 0.23 mV	
	(10 to 30) kHz	0.49 mV/V + 0.23 mV	
	(30 to 50) kHz	0.49 mV/V + 0.23 mV	
	(50 to 100) kHz	3 mV/V + 0.31 mV	
	(100 to 200) kHz	7.5 mV/V + 0.51 mV	
	200 kHz to 1 MHz	20 mV/V + 2 mV	
	(1 to 2) MHz	37 mV/V + 39 mV	
	(2 to 20) V		
	(1 to 10) Hz	0.92 mV/V + 5.2 mV	
	(10 to 50) Hz	0.6 mV/V + 3.1 mV	
	(50 to 100) Hz	0.35 mV/V + 3.1 mV	
	100 Hz to 2 kHz	0.3 mV/V + 3.1 mV	
	(2 to 10) kHz	0.4 mV/V + 3.1 mV	
	(10 to 30) kHz	0.49 mV/V + 3.3 mV	
	(30 to 50) kHz	0.69 mV/V + 3.2 mV	
	(50 to 100) kHz	0.29 mV/V + 3.3 mV	
	(100 to 200) kHz	7.5 mV/V + 5.1 mV	
	200 kHz to 1 MHz	39.8 mV/V + 45 mV	
	(1 to 2) MHz	70 mV/V + 43 mV	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure ^{1,15}	(20 to 200) V		Comparison to Keithley 2002 8.5 Digit Multimeter
	(1 to 10) Hz	0.92 mV/V + 52 mV	
	(10 to 50) Hz	0.49 mV/V + 31 mV	
	(50 to 100) Hz	0.29 mV/V + 32 mV	
	100 Hz to 2 kHz	0.29 mV/V + 32 mV	
	(2 to 10) kHz	0.39 mV/V + 32 mV	
	(10 to 30) kHz	0.49 mV/V + 33 mV	
	(30 to 50) kHz	0.69 mV/V + 33 mV	
	(50 to 100) kHz	3 mV/V + 30 mV	
	(100 to 200) kHz	7.5 mV/V + 53 mV	
	200 kHz to 1 MHz	40 mV/V + 0.4 V	
	(200 to 750) V		
	(1 to 10) Hz	1.3 mV/V + 0.12 V	
	(10 to 50) Hz	0.89 mV/V + 0.12 V	
	(50 to 100) Hz	0.49 mV/V + 0.12 V	
	100 Hz to 2 kHz	0.49 mV/V + 0.12 V	
	(2 to 10) kHz	0.6 mV/V + 0.11 V	
	(10 to 30) kHz	0.57 mV/V + 0.42 V	
	(30 to 50) kHz	0.76 mV/V + 0.4 V	
	(50 to 100) kHz	4.9 mV/V + 0.23 V	
AC Voltage – Measure ^{16,18}	(2 to 200) mV		Comparison to Fluke 8508A 8.5 Digit Multimeter
	(10 to 40) Hz	0.19 μ V/mV + 27 μ V	
	(40 to 100) Hz	0.16 μ V/mV + 3.4 μ V	
	100 Hz to 2 kHz	0.16 μ V/mV + 1.4 μ V	
	(2 to 10) kHz	0.31 μ V/mV + 66 nV	
	(10 to 30) kHz	0.3 μ V/mV + 7.5 μ V	
	(30 to 100) kHz	0.76 μ V/mV + 17 μ V	
	(0.2 to 2 V)		
	(10 to 40) Hz	0.16 mV/V + 0.11 mV	
	(40 to 100) Hz	0.12 mV/V + 15 μ V	
	100 Hz to 2 kHz	76 μ V/V + 17 μ V	
	(2 to 10) kHz	94 μ V/V + 17 μ V	
	(10 to 30) kHz	0.22 mV/V + 38 μ V	
	(30 to 100) kHz	0.47 mV/V + 0.19 mV	
	(100 to 300) kHz	2.4 mV/V + 1.9 mV	
	300 kHz to 1 MHz	7.8 mV/V + 19 mV	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
AC Voltage – Measure ^{16,18}	(2 to 20) V		Comparison to Fluke 8508A 8.5 Digit Multimeter
	(10 to 40) Hz	0.16 mV/V + 0.95 mV	
	(40 to 100) Hz	82 μ V/V + 0.18 mV	
	100 Hz to 2 kHz	62 μ V/V + 0.19 mV	
	(2 to 10) kHz	92 μ V/V + 0.17 mV	
	(10 to 30) kHz	0.22 mV/V + 0.37 mV	
	(30 to 100) kHz	0.48 mV/V + 1.8 mV	
	(100 to 300) kHz	2.4 mV/V + 19 mV	
	300 kHz to 1 MHz	7.8 mV/V + 0.19 V	
	(2 to 200) V		
	(10 to 40) Hz	0.15 mV/V + 9.7 mV	
	(40 to 100) Hz	87 μ V/V + 1.8 mV	
	100 Hz to 2 kHz	74 μ V/V + 1.7 mV	
	(2 to 10) kHz	93 μ V/V + 1.7 mV	
	(10 to 30) kHz	0.22 mV/V + 3.6 mV	
	(30 to 100) kHz	0.48 mV/V + 19 mV	
AC High Voltage – Measure ^{1,15,18}	(200 to 1 000) V		Comparison to Vitretek 4700 High Voltage Meter, HVL-35 Probe, HVL-100 Probe, HVL-150 Probe
	40 Hz to 10 kHz	0.12 mV/V + 19 mV	
	(10 to 30) kHz	0.23 mV/V + 39 mV	
	(30 to 200) Hz		
	(10 to 1 000) V	1.4 mV/V + 0.13 V	
	(1 to 10) kV	1.5 V/kV + 0.4 V	
	(50 to 70) Hz		
	(1 to 20) kV	0.74 V/kV + 0.45 V	
	(20 to 30) kV	0.77 V/kV – 0.28 V	
	(30 or 70) Hz		
	(10 to 50) kV	1.8 V/kV – 3.1 V	
	(50 to 75) kV	2 V/kV – 9.7 V	
	(30 to 65) Hz		
	(10 to 100) kV	6.3 V/kV – 3 V	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Source/Measure ^{1,15}	Type B		Comparison to Fluke 5522A/SC1100 Multi Product Calibrator
	(600 to 800) °C	0.34 °C	
	(800 to 1 000) °C	0.28 °C	
	(1 000 to 1 550) °C	0.25 °C	
	(1 550 to 1 820) °C	0.26 °C	
	Type C		
	(0 to 150) °C	0.23 °C	
	(150 to 650) °C	0.21 °C	
	(650 to 1 000) °C	0.24 °C	
	(1 000 to 1 800) °C	0.39 °C	
	(1 800 to 2 316) °C	0.65 °C	
	Type E		
	(-250 to -100) °C	0.41 °C	
	(-100 to -25) °C	0.13 °C	
	(-25 to 350) °C	0.11 °C	
	(350 to 650) °C	0.13 °C	
	(650 to 1 000) °C	0.17 °C	
	Type J		
	(-210 to -100) °C	0.23 °C	
	(-100 to -30) °C	0.13 °C	
	(-30 to 150) °C	0.11 °C	
	(150 to 760) °C	0.13 °C	
	(760 to 1 200) °C	0.18 °C	
	Type K		
	(-200 to -100) °C	0.26 °C	
	(-100 to -25) °C	0.14 °C	
	(-25 to 120) °C	0.13 °C	
	(120 to 1 000) °C	0.2 °C	
	(1 000 to 1 372) °C	0.31 °C	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Electrical Simulation of Thermocouple Indicating Devices – Source/Measure ^{1,15}	Type N		Comparison to Fluke 5522A/SC1100 Multi Product Calibrator
	(-200 to -100) °C	0.31 °C	
	(-100 to -25) °C	0.17 °C	
	(-25 to 120) °C	0.15 °C	
	(120 to 410) °C	0.14 °C	
	(410 to 1 300) °C	0.21 °C	
	Type R		
	(0 to 250) °C	0.45 °C	
	(250 to 400) °C	0.27 °C	
	(400 to 1 000) °C	0.26 °C	
	(1 000 to 1 767) °C	0.31 °C	
	Type S		
	(0 to 250) °C	0.37 °C	
	(250 to 1 000) °C	0.28 °C	
	(1 000 to 1 400) °C	0.29 °C	
	(1 400 to 1 767) °C	0.36 °C	
	Type T		
	(-250 to -150) °C	0.49 °C	
	(-150 to 0) °C	0.19 °C	
	(0 to 120) °C	0.13 °C	
	(120 to 400) °C	0.11 °C	
Oscilloscopes ^{1,15}			Comparison to Fluke 5522A/SC1100 Multi Product Calibrator
Amplitude – DC Signal into 50 Ω load	± 1 mV to ± 6.6 V	1.5 mV/V + 28 μV	
into 1 MΩ load	± 1 mV to ± 130 V	0.39 mV/V + 32 μV	
Amplitude – Square Wave into 50 Ω load	1 mVp-p to 6.6 Vp-p 10 Hz to 10 kHz	1.9 mV/V + 32 μV	
into 1 MΩ load	1 mVp-p to 130 Vp-p 10 Hz to 1 kHz (1 to 10) kHz	0.78 mV/V + 34 μV 2 mV/V + 48 μV	

Electrical – DC/Low Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Oscilloscopes ^{1,15} Leveled Sine Wave – Amplitude (50 kHz ref) into 50 Ω load	5 mV to 5.5 V 50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz 5 mV to 3.5 V (600 to 1 100) MHz 5 mVp-p to 5.5 Vp-p 50 kHz to 100 MHz (100 to 300) MHz (300 to 600) MHz 4 mVp-p to 3.5 Vp-p (600 to 1 100) MHz	27 mV/V + 1.7 mV 31 mV/V + 0.25 mV 47 mV/V + 0.25 mV 0.42 mV + 54 mV/V 12 mV/V + 0.1 mV 16 mV/V + 0.1 mV 31 mV/V + 0.1 mV 39 mV/V + 90 μ V	Comparison to Fluke 5522A/SC1100 Multi Product Calibrator
Time Markers	1 ns to 20 ms 50 ms 0.1 s 0.2 s 0.5 s 1 s 2 s 5 s	2. ps/ μ s 4.4 μ s 9.8 μ s 35 μ s 0.2 ms 0.8 ms 3.1 ms 20 ms	
Edge – Rise Time	(250 to 350) ps 1 kHz to 11 MHz	310 ps	
Total Harmonic Distortion ^{1,15}	10 Hz to 500 kHz (-95 to -50) dB (-50 to -20) dB (-20 to 0) dB	0.021 dB 0.13 dB 0.89 dB	Comparison to Pico Technologies 4262 Digital Oscilloscope

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Caliper Masters ²	(0.5 to 4) in (4 to 40) in (40 to 60) in	(1.8 + 4L) μin (4.5L) μin (130 + 1L) μin	Comparison to P&W Labmaster® Universal 1000A, Gauge Blocks
Cylindrical Plug Gauges ^{2,12}	(0.01 to 0.1) in (0.1 to 0.42) in (0.42 to 4) in (4 to 12) in	(3.1 – 2L) μin (2.9 + 2L) μin (1.9 + 4L) μin (1 + 4L) μin	Comparison to P&W Labmaster® Universal 1000A, Gauge Blocks
Cylindrical Ring Gauges ^{2,14}	(0.01 to 0.22) in (0.22 to 1) in (1 to 4) in (4 to 7) in (7 to 10) in (10 to 13) in	(8.3 + 2.8L) μin (8 + 4.5L) μin (10 + 2.7L) μin (22 + 2.7L) μin (9.2 + 5L) μin (9.4 + 5L) μin	Comparison to P&W Labmaster® Universal 1000A, Gauge Blocks, Master Ring Gauges
Depth Micrometer Master ²	(0.5 to 11.5) in	(28 + 1L) μin	Comparison to Gauging Amplifier, Gauge Blocks
End Measuring Rods ²	(0.5 to 4) in (4 to 20) in (20 to 80) in	(4.1 + 3L) μin (0.61 + 4L) μin (100 + 4L) μin	Comparison to P&W Labmaster® Universal 1000A, Gauge Blocks
Feeler Gauges ¹ (Leaf-Style)	Up to 0.25 in	76 μin	Comparison to Bench Micrometer
Gauge Balls ² (size only)	(0.062 5 to 0.1) in (0.1 to 0.42) in (0.42 to 2) in	(10 – 2L) μin (9 + 2L) μin (8 + 5L) μin	Comparison to P&W Labmaster® Universal 1000A, Gauge Blocks
Gauge Blocks ^{2,6}	(0.005 to 4) in	(3 + 0.45L) μin	Comparison to Mahr 130-B24 Comparator, Master Gauge Blocks
Gauge Blocks ^{2,6,18}	(0.005 to 0.1) in (0.1 to 0.4) in (0.4 to 4) in (5 to 20) in	(3.2 + 6.2L) μin (2.9 + 3L) μin (3.7 + 0.97L) μin (1.6 + 1.4L) μin	Comparison to P&W Labmaster® Universal 1000A, Master Gauge Blocks
Micrometer Masters ²	Up to 0.1 in (0.1 to 0.42) in (0.42 to 4) in (4 to 20) in (20 to 40) in	(3.1 – 4L) μin (2.6 + 2L) μin (1.8 + 4L) μin (0.53 + 4L) μin (4L – 8.1) μin	Comparison to P&W Labmaster® Universal 1000A, Gauge Blocks

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Optical Flats and Parallels Flatness	Up to 4 in Diameter Up to 100 μin	5.6 μin	Comparison to Optical Flat, Monochromatic Light
Parallelism	(0 to 2) in Thickness	4.5 μin	Mahr 130-B24 Comparator
Parallels	Up to 36 in	66 μin	Comparison to Gauging Amplifier, Gauge Blocks
Pin Gauges – Class ZZ ²	(0.01 to 1) in	(93 + 2.4L) μin	Comparison to Digital Micrometer
Riser Blocks ²	(6 to 24) in	(19 + 7L) μin	Comparison to Gauging Amplifier, Gauge Blocks
Snap Gauges ²	(0.01 to 0.22) in (0.22 to 1) in (1 to 13) in	(23 – 1L) μin (24 – 4L) μin (16 + 3.7L) μin	Comparison to P&W Labmaster® Universal 1000A, Gauge Blocks
Squares – Perpendicularity ²	Up to 36 in	(35 + 8L) μin	Comparison to Indi-Square, Gauging Amplifier, Gauge Blocks, Tri-Square
Tapered Plugs ²	Up to 0.1 in (0.1 to 0.42) in (0.42 to 2) in	(23 – 0.8L) μin (22 + 0.26L) μin (22 + 0.87L) μin	Comparison to P&W Labmaster® Universal 1000A, Gauge Blocks, Plug Gauges
Tapered Rings ²	(0.01 to 0.22) in (0.22 to 1) in (1 to 2) in	(24 – 0.14L) μin (25 – 1.7L) μin (17 + 6.3L) μin	Comparison to P&W Labmaster® Universal 1000A, Gauge Blocks
Thickness (Film) Gauge Standards (Non-Ferrous)	Up to 0.05 in	6.1 μin	Comparison to P&W Labmaster® Universal 1000A, Gauge Blocks
Thread Measuring Wires ^{2,10}	Unified 60° (4 to 80) TPI Acme 29° (1 to 20) TPI	(13 – 50L) μin (13 – 50L) μin	Comparison to P&W Labmaster® Universal 1000A, Gauge Blocks

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Thread Micrometer Standards	1 in 2 in 3 in 4 in 5 in 6 in	8.8 µin 16 µin 24 µin 31 µin 40 µin 48 µin	Comparison to P&W Labmaster® Universal 1000A, Gauge Blocks
Angle Blocks	(1 to 60)°	0.002 2°	Comparison to Master Angle Blocks, Sine Plate, Gauging Amplifier
Angle Gauges (Leaf Style) ²	Up to 90°	3.7'	Comparison to STI Optical Comparator
Electronic Differential Levels ²	Up to 1 000"	1.4"	Comparison to Brunson 470 Angle Generator
Functional Gauges & Fixtures Linear	Up to 12 in	190 µin	Comparison to STI Optical Comparator
Angle	Up to 90°	0.065°	
Radius Gauges (Leaf Style)	Up to 1 in	210 µin	Comparison to STI Optical Comparator
Sine Bars, Sine Plates ²	Angle (1 to 60)°	6.5"	Comparison to Gauge Blocks, Angle Blocks, Gauging Amplifier
Thread Pitch Gauges ¹⁰ (Leaf Style)	(4 to 84) TPI	160 µin	Comparison to STI Optical Comparator
Tri-Blocks			
Length	(1 to 6) in	48 µin	Comparisons to Gauging Amplifier, Gauge Blocks
Flatness	(1 to 6) in	41 µin	Gauging Amplifier
Perpendicularity	(1 to 6) in	69 µin	Indi-Square, Gauging Amplifier

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
V-Blocks			
Parallelism	Up to 3 in	47 μ in	Comparisons to Gauging Amplifier, Gauge Blocks
Perpendicularity	Up to 6 in	66 μ in	Gauging Amplifier, Indi-Square
V-Centrality	Up to 3 in	55 μ in	Gauging Amplifier, Master Plugs
Pipe Thread Plugs ^{2,10}			
Simple Pitch Diameter	(4 to 80 TPI)	(100 + 2L) μ in	Comparisons to P&W Labmaster® Universal 1000A, Gauge Blocks, Sine Plug
Standoff	Up to 1 in (1 to 2) in	(36 + 0.9L) μ in (35 + 0.59L) μ in	Gauging Amplifier, Gauge Blocks
Thread Plugs – Setting ^{2,5,10}			
Simple Pitch Diameter	(4 to 80 TPI)	(47 + 5L) μ in	Comparisons to P&W Labmaster® Universal 1000A, Gauge Blocks, Thread Measuring Wires
Major Diameter	(0.06 to 4) in	(16 + 4.7L) μ in	P&W Labmaster® Universal 1000A, Gauge Blocks
Root Radius and Minor Diameter	(0.06 to 4) in	203 μ in	STI Optical Comparator
Thread Plugs – Working ^{2,5,10}			
Simple Pitch Diameter	(4 to 80 TPI)	(96 + 0.4L) μ in	Comparisons to P&W Labmaster® Universal 1000A, Gauge Blocks, Thread Measuring Wires
Major Diameter	(0.06 to 4) in	(26 + 4L) μ in	P&W Labmaster® Universal 1000A, Gauge Blocks

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Thread Plugs – Working ^{2,5,10} Root Radius and Minor Diameter	(0.06 to 4) in	200 μ in	Comparison to STI Optical Comparator
Bench Micrometers ¹ Length	Up to 2 in	12 μ in	Comparisons to Gauge Blocks
Anvil Flatness	Up to 0.5 in Diameter	9.4 μ in	Optical Flat, Monochromatic Light
Bore Gauges ² (2 point)	Up to 4 in (4 to 40) in	(8.4 + 7L) μ in (3.6 + 8L) μ in	Comparison to P&W Labmaster® Universal 1000A, Gauge Blocks
Bore Gauges ² (3 point)	Up to 8 in	(72 + 3L) μ in	Comparison to Ring Gauges
Calipers ^{1,2,3,18} (Inside and Outside)	Up to 5 in (5 to 20) in	(3.4 + 7.1L) μ in (2.6 + 7.9L) μ in	Comparison to Gauge Blocks, Gauge Block Accessories
Calipers ^{1,2,3} (Inside and Outside)	(20 to 40) in (40 to 60) in	(440 + 5L) μ in (920 + 4L) μ in	Comparison to Gauge Blocks, Gauge Block Accessories
Chamfer Gauges ^{1,2}	Up to 1 in	(77 + 3.2L) μ in	Comparison to Modified Ring Gauges
Gauging Amplifiers, LVDT Heads	Up to 0.001 in	(6.8 – 3L) μ in	Comparison to P&W Labmaster® Universal 1000A
Height Gauges ^{1,2}	Up to 24 in (24 to 40) in	(430 + 2L) μ in (390 + 4L) μ in	Comparison to Gauge Blocks, Surface Plate
Height Masters ² Length	Up to 24 in	(68 + 6L) μ in	Comparisons to Gauging Amplifier, Gauge Blocks
Parallelism	Up to 0.001 in	15 μ in	
Indicators ^{2,7}	(0 to 0.1) in (0.1 to 0.42) in (0.42 to 4) in (4 to 8) in	(58 – 0.3L) μ in (58 + 0.1L) μ in (58 + 0.6L) μ in (46 + 4L) μ in	Comparison to P&W Labmaster® Universal 1000A, Gauge Blocks

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Micrometer Heads ² Length	(0 to 0.1) in (0.1 to 0.42) in (0.42 to 2) in	$(11 - 2L) \mu\text{in}$ $(10 + 1.6L) \mu\text{in}$ $(8.9 + 4.72L) \mu\text{in}$	Comparisons to P&W Labmaster® Universal 1000A, Gauge Blocks
Anvil Flatness	Up to 3 in Diameter	9.4 μin	Optical Flat, Monochromatic Light
Depth Micrometers ^{1,2} Length	Up to 12 in	$(45 + 5L) \mu\text{in}$	Comparisons to Gauge Blocks
Base Flatness	Up to 3 in	9.4 μin	Optical Flat, Monochromatic Light
Inside Micrometer ^{1,2}	Up to 4 in (4 to 20) in (21 to 40) in (41 to 60) in	$(37 + 5L) \mu\text{in}$ $(55 + 7L) \mu\text{in}$ $(300 + 6L) \mu\text{in}$ $(590 + 4L) \mu\text{in}$	Comparison to Gauge Blocks, Gauge Block Accessories
Outside Micrometer ^{1,2}	Up to 4 in (4 to 20) in (21 to 40) in (41 to 60) in	$(37 + 5L) \mu\text{in}$ $(55 + 7L) \mu\text{in}$ $(270 + 6L) \mu\text{in}$ $(580 + 5L) \mu\text{in}$	Comparison to Gauge Blocks
Micrometer Anvil Flatness	Up to 3 in Diameter	9.4 μin	Comparison to Optical Flat, Monochromatic Light
Screw Thread Micrometer ^{2,4}	Up to 1 in	$(160 + 7.7L) \mu\text{in}$	Comparison to Gage Blocks, STI Optical Comparator
V-Anvil Micrometer ^{1,2}	(0.062 5 to 2) in	$(53 + 1L) \mu\text{in}$	Comparison to Gauge Balls
Steel Rules, Tape Measures ^{2,11}	Up to 12 in	890 μin	Comparison to Optical Comparator
Steel Rules, Tape Measures ^{2,11}	(1 to 300) ft	0.036 in	Comparison to Optodyne LDDM
Thickness Gauges ¹ (Dial & Digital)	Up to 0.5 in	60 μin	Comparison to Gauge Blocks

Length – Dimensional Metrology

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Optical Comparators, Vision Measuring Machines ¹ Magnification	10X, 20X, 31.25X, 50X and 62.5X	0.001 2 in	Comparisons to Magnification Checker Glass Scale
Linear Length	X and Y Up to 6 in (6 to 12) in	110 μ in 150 μ in	Glass Scale Gauge Blocks
Angle	Up to 90°	0.021°	Angle Blocks
Protractors, Inclinometers ²	Up to 90°	0.59'	Comparison to Gauge Blocks, Angle Blocks, Gauging Amplifier, Sine Plate
Levels ²	Up to 1 000"	2.9"	Comparison to Brunson 470 Angle Generator

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Force Gauges (Tension & Compression)	(0.5 to 5) lbf (5 to 50) lbf (50 to 600) lbf	0.2 % of reading + 0.000 2 lbf 0.02 % of reading + 0.008 lbf 0.02 % of reading + 0.042 lbf	Comparison to NIST Class F Weights
Force Gauges ¹⁸ (Tension & Compression)	(2.5 to 100) lbf (20 to 1 000) lbf	0.000 16 % of reading + 0.01 lbf 0.000 18 % of reading + 0.1 lbf	Comparison to Morehouse Force Calibration System, Load Cell
Cable/Wire Tensiometers	(1 to 600) lbf	0.01 % of reading + 0.058 lbf	Comparison to NIST Class F Weights

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Durometers			
Indenter Dimensions			Full Verification per ASTM D2240 using STI Optical Comparator
Angle	Up to 35°	0.065°	
Radius	Up to 0.05 in	250 μin	
Diameter/Length	Up to 0.2 in	180 μin	
Spring Force			Durometer Calibrator
Types A, B, E, O	Up to 7.3 N	0.05 N	
Types C, D, DO	Up to 40 N	0.28 N	
Durometer Calibrator Force			Comparison to ASTM E617 Class 1 Weights
Types A, B, E, O	Up to 7.3 N	0.000 23 N	
Types C, D, DO	Up to 40 N	0.000 85 N	
Hydraulic Pressure – Measure/Source	(10 to 50) psig (50 to 500) psig (500 to 10 000) psig	0.011 psi 0.035 psi 0.007 9 % of reading + 0.043 psi	Comparison to Fluke P3224-PSI Deadweight Tester
Low Pressure/Vacuum – Measure/Source	Up to 2 inH ₂ O	0.000 69 inH ₂ O	Comparison to Dwyer 1430 Microtector
Low Pressure/Vacuum – Measure/Source	(-20 to 20) inH ₂ O	0.003 1 inH ₂ O	Comparison to Meriam 34FBT2M Manometer
Pneumatic Absolute Pressure – Measure/Source	(0.2 to 1 015) psia	0.002 3 % of reading	Comparison to Ruska 2465 Deadweight Tester
Pneumatic Gauge Pressure – Measure/Source	(-14.4 to 1 000) psig	0.002 3 % of reading	Comparison to Ruska 2465 Deadweight Tester
Pressure/Vacuum – Measure/Source ^{1,15}	(0 to 15) psig (-14.4 to 0) psiv (-14.4 to 30) psig Up to 100 psig Up to 500 psig Up to 1 000 psig Up to 10 000 psig	0.008 9 psi 0.011 psi 0.023 psi 0.056 psi 0.25 psi 0.59 psi 8.3 psi	Comparison to Fluke 700 Series Pressure Transducers
Scales & Balances ¹ (0.000 1 g resolution)	Up to 610 g	0.000 2 % of reading + 0.1 mg	ASTM E617 Class 1 weights and NIST HB 44 utilized in the calibration of the weighing system.

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Scales & Balances ¹ (0.001 g resolution)	Up to 5 000 g	0.000 2 % of reading + 1 mg	ASTM E617 Class 1 weights and NIST HB 44 utilized in the calibration of the weighing system.
Scales & Balances ¹ (0.01 g resolution)	Up to 35 000 g	0.000 2 % of reading + 10 mg	ASTM E617 Class 1 weights and NIST HB 44 utilized in the calibration of the weighing system.
Scales & Balances ¹ (0.000 1 kg resolution)	Up to 35 kg	0.000 8 % of reading + 95 mg	ASTM E617 Class 1 weights and NIST HB 44 utilized in the calibration of the weighing system.
Scales & Balances ¹ (0.001 lb resolution)	Up to 10 lb	0.004 % of reading + 0.001 lb	NIST Class F weights and NIST HB 44 utilized in the calibration of the weighing system.
Scales & Balances ¹ (0.01 lb resolution)	Up to 100 lb	0.005 % of reading + 0.009 5 lb	NIST Class F weights and NIST HB 44 utilized in the calibration of the weighing system.
Scales & Balances ¹ (0.1 lb resolution)	Up to 600 lb	0.003 % of reading + 0.1 lb	NIST Class F weights and NIST HB 44 utilized in the calibration of the weighing system.
Torque Indicating Devices	(20 to 200) ozf·in (4 to 50) lbf·in (30 to 400) lbf·in (80 to 1000) lbf·in (20 to 250) lbf·ft (100 to 1 000) lbf·ft (200 to 2 000) lbf·ft	0.2 % of reading + 0.05 ozf·in 0.2 % of reading + 0.009 lbf·in 0.2 % of reading + 0.06 lbf·in 0.18 % of reading + 0.98 lbf·in 0.14 % of reading + 0.4 lbf·ft 0.61 % of reading + 0.01 lbf·ft 0.5 % of reading + 0.02 lbf·ft	Comparison to CDI 2000 Torque Tester
Torque Calibration Systems	(20 to 400) ozf·in (2.5 to 100) lbf·in (40 to 3 000) lbf·in (200 to 2 000) lbf·ft	0.2 % of reading + 0.002 ozf·in 0.06 % of reading + 0.009 lbf·in 0.038 % of reading + 0.19 lbf·in 0.04% of reading + 0.09 lbf·ft	Comparison to CDI Torque Arms, CDI Torque Wheels, NIST Class F Weights

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Humidity – Measure ^{1,15,18}	(10 to 30) °C (10 to 90) %RH	0.12 % of reading + 1.5 %RH	Comparison to Rotronic HC2A Humidity Probe
Temperature and Humidity Devices ¹⁸			
Humidity	(15 to 40) °C (5 to 95) %RH	0.27 % of reading + 0.56 %RH	Comparison to Thunder Scientific 2500LT Two-pressure Humidity Generation System
Temperature	(-10 to 70) °C	0.000 51 % of reading + 0.14 °C	
Dew Point	(-25 to 69) °C	0.22 °C	Comparison to Thunder Scientific 2500LT Chamber
Radiation (Infrared) Thermometers ^{1,15}	50 °C 100 °C 200 °C 300 °C 400 °C 500 °C	0.86 °C 1.6 °C 3.1 °C 4.7 °C 6.4 °C 8.2 °C	Comparison to Fluke 9132 Blackbody Source (Flat Plate) $\epsilon = 0.95$, $\lambda = (8 \text{ to } 14) \mu\text{m}$
Radiation (Infrared) Thermometers	-15 °C 0 °C 50 °C 100 °C 120 °C	0.78 °C 0.57 °C 0.7 °C 0.69 °C 0.75 °C	Comparison to Fluke 4180 Blackbody Source (Flat Plate) $\epsilon = (0.9 \text{ to } 1)$, $\lambda = (8 \text{ to } 14) \mu\text{m}$
Radiation (Infrared) Thermometers	35 °C 100 °C 200 °C 350 °C 500 °C	0.54 °C 0.84 °C 1 °C 1.8 °C 2.4 °C	Comparison to Fluke 4181 Blackbody Source (Flat Plate) $\epsilon = (0.9 \text{ to } 1)$, $\lambda = (8 \text{ to } 14) \mu\text{m}$
Temperature – Source ⁸	-196 °C	0.028 °C	Comparison to Fluke 7196 Liquid N ₂ Calibrator, Fluke 5628 PRT
Temperature – Source ^{1,8,15}	0 °C	0.058 °C	Comparison to Kaye X0240 Ice Point
Temperature – Source ⁸	0.01 °C	0.000 48 °C	Comparison to Triple Point of Water Cell

Thermodynamic

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Temperature – Source	0 °C 25 °C 50 °C (-50 to 0) °C (0 to 100) °C	0.028 °C 0.025 °C 0.03 °C 0.056 °C 0.057 °C	Comparison to Fluidized Baths, Thermometrics 1925-A, Fluke 5628 PRT
Temperature – Source ⁸	(100 to 250) °C	0.06 % of reading	Comparison to Fluidized Baths, Thermometrics 1925-A, Fluke 5628 PRT
Temperature – Source ^{1,8,15,18}	(-15 to 110) °C (50 to 350) °C	0.000 18 % of reading + 0.3 °C 0.000 12 % of reading + 0.72 °C	Comparison to Hart 9009 Drywell
Temperature – Source ⁸	(-95 to 140) °C	0.032 °C	Comparison to Fluke 9190A Drywell, Fluke 5628 PRT
Temperature – Source ⁸	(-45 to 140) °C	0.045 °C	Comparison to Fluke 9170 Drywell, Fluke 5628 PRT
Temperature – Source ⁸	(50 to 700) °C	0.003 7 % of reading + 0.087 °C	Comparison to Fluke 9173 Drywell, Fluke 5628 PRT
Temperature – Measure	(-200 to 660) °C	0.004 % of reading + 0.013 °C	Comparison to Fluke 1586 Scanner, Fluke 5628 PRT
Temperature – Measure	(0 to 100) °C	0.003 % of reading + 0.009 °C	Comparison to Fluke 1586 Scanner, Thermometrics 1925-A
Temperature – Measure ^{1,15,18}	(-10 to 60) °C	0.4 °C	Comparison to Rotronic HygroPalm HP22-A, HC2A
Temperature – Measure ^{1,15,18}	(-40 to 0) °C (0 to 232) °C	0.7 °C 0.03 % of reading + 0.7 °C	Comparison to Type T Thermocouple Probe, Fluke 52 II Temperature Indicator
Temperature – Measure ^{1,15,18}	(0 to 232) °C	1.5 °C	Comparison to Type K Thermocouple Probe, Fluke 52 II Temperature Indicator

Time and Frequency

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Frequency – Measure	10 Hz to 1.3 GHz	5.3 pHz/Hz + 0.1 mHz	Comparison to Keysight 53220A Counter, HP 58503A GPS Receiver
Frequency – Measure ^{1,15}	0.1 Hz to 160 MHz	29 nHz/Hz + 6.1 mHz	Comparison to Phillips PM6669 Frequency Counter, Efratom FRK-LN Oscillator
Frequency – Source	10 Hz to 1.3 GHz	59 pHz/Hz + 0.1 mHz	Comparison to Keysight 53220A Universal Counter, Efratom FRK-LN Oscillator
Non-contact Tachometers ^{1,2,15,18}	(15 to 90 000) rpm	0.000 58 % of reading	Comparison to Fluke 5522A Multi Product Calibrator, Light Emitting Diode
Strobe Tachometers ^{1,2,15}	(25 to 90 000) rpm	0.000 5 % of reading + 0.13 rpm	Comparison to Keysight 53220A Counter, Solar Cell
Stopwatches & Timers ^{15,18}	Up to 24 h	46 ms/24 h	Comparison to Helmut Klein 4500 Timometer

DIMENSIONAL MEASUREMENT

1 Dimensional

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Length Measures ²	Up to 300 in	$(3\ 100 + 10.2L)\ \mu\text{in}$	1D Measurements performed using ProScale Optodyne LDDM

1 Dimensional

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Length Measures – External ^{2,9}	Up to 0.1 in (0.1 to 0.42) in (0.5 to 4) in (4 to 20) in (20 to 80) in	$(3.2 - 3L) \mu\text{in}$ $(2.3 + 5L) \mu\text{in}$ $(1.6 + 4L) \mu\text{in}$ $(0.61 + 4L) \mu\text{in}$ $(75 + 5L) \mu\text{in}$	1D Measurements performed using P&W Labmaster® Universal 1000A P&W Labmaster Universal 1000A
Length Measures – Internal ^{2,9}	(0.01 to 0.22) in (0.22 to 1) in (1 to 2) in (2 to 3) in (3 to 4) in (4 to 7) in (7 to 10) in (10 to 13) in	$(8.8 + 0.7L) \mu\text{in}$ $(8 + 4.5L) \mu\text{in}$ $(7.9 + 5L) \mu\text{in}$ $(8.1 + 5L) \mu\text{in}$ $(8.4 + 5L) \mu\text{in}$ $(8.7 + 5L) \mu\text{in}$ $(9.2 + 5L) \mu\text{in}$ $(9.9 + 5L) \mu\text{in}$	1D Measurements performed using P&W Labmaster® Universal 1000A P&W Labmaster Universal 1000A
Length Measures ^{2,9}	Up to 1 in	$(93 + 2.44L) \mu\text{in}$	1D Measurements performed using Digital Micrometer
Length Measures ^{2,9}	Up to 8 in	$(1\,400 + 5L) \mu\text{in}$	1D Measurements performed using Digital Caliper

2 Dimensional

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
2D Angles	Up to 360°	0.065°	2D Measurements performed using STI Optical Comparator
2D Length ⁹	X/Y Axis: Up to 12 in	210 μin	2D Measurements performed using STI Optical Comparator

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

- On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
- L = length in inches, DL = Diagonal Length; ' = arc-minute; " = arc-second; rpm = revolutions per minute.
- Measurements include the following measurement functions: Outside, Inside, Step and Depth Extension Rods.
- Inch thread setting plugs only with 60° Included Angle.
- Includes 60° Metric, Unified and 55° Whitworth pitch gauges.

6. Uncertainty is for Steel Blocks. Carbide and Ceramic Blocks may have a different uncertainty due to deformation coefficients and different coefficients of thermal expansion.
7. Includes dial, digital and test indicators.
8. Includes Liquid-in-Glass Thermometers, RTDs, Thermocouples, Bi-metallic Thermometers, etc. Liquid-in-Glass Thermometers are only calibrated in fluidized baths to ensure correct immersion depth and stem effect corrections.
9. Metric equivalencies for this type of equipment are available and converted by 1 in equals 25.4 mm exactly.
10. TPI indicates threads per inch.
11. Verification performed in 12 ft. increments before repositioning.
12. Includes Master Setting Discs and Progressive Diameter Plugs.
13. This instrument/parameter has been characterized to lower the uncertainty.
14. The stated uncertainty is associated with a primary calibration which utilizes a comparison datum set with gauge blocks rather than a master ring.
15. On-site calibration service is available for this parameter based out of the Brookfield, Wisconsin office.. Since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope.
16. Above 300 V, < 10 kHz, add $0.0004(R-300)^2$ ppm; > 10 kHz, add $(0.0004 + (F-10000)*1E-07)*(R-300)^2$ ppm.
17. Power uncertainties are estimates based on generalized voltage and current outputs. Actual uncertainties may differ due to actual voltage and/or current values used to simulate power. Power uncertainties are calculated by combining the voltage, current, and phase uncertainties via the root sum square method.
18. The uncertainty contributed by the resolution of the Unit Under Test is accounted for at the time of the calibration and is not included in the CMC.
19. Unless otherwise specified in the far-right column, the calibration procedure/method utilized by the laboratory was developed and validated internally.



Jason Stine, Vice President

